

Manufacturer name:	Medline															
About the Recall: <i>Clear overview of recall and products affected</i>	An Urgent Product Recall regarding specific lot(s) of their Leg Bags. This notice has been issued due to the discovery of Latex material being used in the leg strap component of the leg beg during testing. Product labeling states “Not made with Natural Rubber Latex”.															
Please list the specific products(s) numbers/models that are affected:	<table border="1"> <thead> <tr> <th data-bbox="472 579 573 625">MMS #</th> <th data-bbox="573 579 753 625">MFG Catalog #</th> <th data-bbox="753 579 1224 625">Description</th> <th data-bbox="1224 579 1409 625">Affected Lot(s)</th> </tr> </thead> <tbody> <tr> <td data-bbox="472 625 573 753">476196</td> <td data-bbox="573 625 753 753">DYND12578</td> <td data-bbox="753 625 1224 753">BAG, LEG W/TWIST VLV LF LG 32OZ (48/CS)</td> <td data-bbox="1224 625 1409 753">48623040001 48623050001 48623060001 48623070001 48623080001</td> </tr> <tr> <td data-bbox="472 753 573 861">446577</td> <td data-bbox="573 753 753 861">DYND12574</td> <td data-bbox="753 753 1224 861">LEGBAG, W/STRAP (48/CS)</td> <td data-bbox="1224 753 1409 861">48623040001 48623050001 48623060001 48623080001</td> </tr> </tbody> </table>				MMS #	MFG Catalog #	Description	Affected Lot(s)	476196	DYND12578	BAG, LEG W/TWIST VLV LF LG 32OZ (48/CS)	48623040001 48623050001 48623060001 48623070001 48623080001	446577	DYND12574	LEGBAG, W/STRAP (48/CS)	48623040001 48623050001 48623060001 48623080001
MMS #	MFG Catalog #	Description	Affected Lot(s)													
476196	DYND12578	BAG, LEG W/TWIST VLV LF LG 32OZ (48/CS)	48623040001 48623050001 48623060001 48623070001 48623080001													
446577	DYND12574	LEGBAG, W/STRAP (48/CS)	48623040001 48623050001 48623060001 48623080001													
URL Link(s) to Manufacturer Notice/Web Site:	www.medline.com															
Actions to be taken by customer/Patient: <i>Please include phone numbers and emails when applicable</i>	Immediately discontinue use of any product matching the affected item(s) and lot number(s) listed above. If you have no product matching the affected item(s) and lot number(s), no further action is needed. Medline: Toll Free: 866.359.1704 Email: Recalls@Medline.com															