

Manufacturer name:	Nurse Assist, LLC																																															
About the Recall: <i>Clear overview of recall and products affected</i>	Urgent Field Safety Notice regarding specific lots of Sterile 0.9% Normal Saline, USP (100 mL bottles, 250 mL bottles, 500 mL bottles, 1000mL bottles, 3.1oz spray can, 7.1oz spray can, 3mL syringes, 5mL syringes, and 10mL syringes) and Sterile Water, USP (100 mL bottles, 250 mL bottles, 500 mL bottles, 1000mL bottles, 120 mL cups, 10mL syringes, and 30mL syringes). This notice has been issued because product cannot be verified as having the required sterility assurance level of (SAL) of 10-6.																																															
Please list the specific products(s) numbers/models that are affected:	<table border="1"> <thead> <tr> <th data-bbox="477 632 574 684">MMS #</th> <th data-bbox="574 632 688 684">MFG Catalog #</th> <th data-bbox="688 632 1019 684">Description</th> <th data-bbox="1019 632 1414 684">Affected Lot(s)</th> </tr> </thead> <tbody> <tr> <td data-bbox="477 684 574 730">560285</td> <td data-bbox="574 684 688 730">37-6210</td> <td data-bbox="688 684 1019 730">WATER, IRR SOL FOIL-LID CUP STR 120ML (48/CS)</td> <td data-bbox="1019 684 1414 1331" rowspan="14" style="text-align: center; vertical-align: middle;"> Expiration Date range from November 1, 2023 to September 18, 2025 </td> </tr> <tr> <td data-bbox="477 730 574 779">560286</td> <td data-bbox="574 730 688 779">37-6220</td> <td data-bbox="688 730 1019 779">SALINE, IRR SOL FOIL-LID CUP STR 120ML (48/CS)</td> </tr> <tr> <td data-bbox="477 779 574 827">560284</td> <td data-bbox="574 779 688 827">37-6240</td> <td data-bbox="688 779 1019 827">SALINE, IRR SOL STR 100ML (48/CS)</td> </tr> <tr> <td data-bbox="477 827 574 875">560283</td> <td data-bbox="574 827 688 875">37-6250</td> <td data-bbox="688 827 1019 875">WATER, IRR SOL STR 100ML (48/CS)</td> </tr> <tr> <td data-bbox="477 875 574 924">520119</td> <td data-bbox="574 875 688 924">37-6260</td> <td data-bbox="688 875 1019 924">WATER, IRR SOL STR 250ML (24/CS)</td> </tr> <tr> <td data-bbox="477 924 574 972">520118</td> <td data-bbox="574 924 688 972">37-6270</td> <td data-bbox="688 924 1019 972">SALINE, IRR SOL STR 250ML (24/CS)</td> </tr> <tr> <td data-bbox="477 972 574 1020">733255</td> <td data-bbox="574 972 688 1020">37-6280</td> <td data-bbox="688 972 1019 1020">SALINE, IRR SOL 0.9% 500ML (18/CS)</td> </tr> <tr> <td data-bbox="477 1020 574 1068">1086157</td> <td data-bbox="574 1020 688 1068">37-6281</td> <td data-bbox="688 1020 1019 1068">SALINE, IRR SOL 0.9% 1000ML (6/CS)</td> </tr> <tr> <td data-bbox="477 1068 574 1117">733254</td> <td data-bbox="574 1068 688 1117">37-6290</td> <td data-bbox="688 1068 1019 1117">WATER, IRR SOL STR 500ML (18/CS)</td> </tr> <tr> <td data-bbox="477 1117 574 1165">1086158</td> <td data-bbox="574 1117 688 1165">37-6291</td> <td data-bbox="688 1117 1019 1165">WATER, IRR SOL STR 1000ML (6/CS)</td> </tr> <tr> <td data-bbox="477 1165 574 1213">1103071</td> <td data-bbox="574 1165 688 1213">37-6503</td> <td data-bbox="688 1165 1019 1213">CLEANSER, WOUND SALINE FLUSH 3.0OZ (1EA/CS 12EA/CS)</td> </tr> <tr> <td data-bbox="477 1213 574 1262">636352</td> <td data-bbox="574 1213 688 1262">37-6507</td> <td data-bbox="688 1213 1019 1262">CLEANSER, WND SALINE FLUSH 7.1OZ (12/CS)</td> </tr> <tr> <td data-bbox="477 1262 574 1331">216197</td> <td data-bbox="574 1262 688 1331">6220</td> <td data-bbox="688 1262 1019 1331">SALINE, IRR SOL STR 120ML (48/CS)</td> </tr> </tbody> </table>				MMS #	MFG Catalog #	Description	Affected Lot(s)	560285	37-6210	WATER, IRR SOL FOIL-LID CUP STR 120ML (48/CS)	Expiration Date range from November 1, 2023 to September 18, 2025	560286	37-6220	SALINE, IRR SOL FOIL-LID CUP STR 120ML (48/CS)	560284	37-6240	SALINE, IRR SOL STR 100ML (48/CS)	560283	37-6250	WATER, IRR SOL STR 100ML (48/CS)	520119	37-6260	WATER, IRR SOL STR 250ML (24/CS)	520118	37-6270	SALINE, IRR SOL STR 250ML (24/CS)	733255	37-6280	SALINE, IRR SOL 0.9% 500ML (18/CS)	1086157	37-6281	SALINE, IRR SOL 0.9% 1000ML (6/CS)	733254	37-6290	WATER, IRR SOL STR 500ML (18/CS)	1086158	37-6291	WATER, IRR SOL STR 1000ML (6/CS)	1103071	37-6503	CLEANSER, WOUND SALINE FLUSH 3.0OZ (1EA/CS 12EA/CS)	636352	37-6507	CLEANSER, WND SALINE FLUSH 7.1OZ (12/CS)	216197	6220	SALINE, IRR SOL STR 120ML (48/CS)
MMS #	MFG Catalog #	Description	Affected Lot(s)																																													
560285	37-6210	WATER, IRR SOL FOIL-LID CUP STR 120ML (48/CS)	Expiration Date range from November 1, 2023 to September 18, 2025																																													
560286	37-6220	SALINE, IRR SOL FOIL-LID CUP STR 120ML (48/CS)																																														
560284	37-6240	SALINE, IRR SOL STR 100ML (48/CS)																																														
560283	37-6250	WATER, IRR SOL STR 100ML (48/CS)																																														
520119	37-6260	WATER, IRR SOL STR 250ML (24/CS)																																														
520118	37-6270	SALINE, IRR SOL STR 250ML (24/CS)																																														
733255	37-6280	SALINE, IRR SOL 0.9% 500ML (18/CS)																																														
1086157	37-6281	SALINE, IRR SOL 0.9% 1000ML (6/CS)																																														
733254	37-6290	WATER, IRR SOL STR 500ML (18/CS)																																														
1086158	37-6291	WATER, IRR SOL STR 1000ML (6/CS)																																														
1103071	37-6503	CLEANSER, WOUND SALINE FLUSH 3.0OZ (1EA/CS 12EA/CS)																																														
636352	37-6507	CLEANSER, WND SALINE FLUSH 7.1OZ (12/CS)																																														
216197	6220	SALINE, IRR SOL STR 120ML (48/CS)																																														
URL Link(s) to Manufacturer Notice/Web Site:	Medical Water Products Sterile Saline Water SteriCare Solutions																																															
Actions to be taken by customer/Patient: <i>Please include phone numbers and emails when applicable</i>	Immediately discontinue use of any product matching the affected item(s) and expiration date(s) listed above. If you have no product matching the affected item(s) and lot number(s), no further action is needed. For questions regarding this notification, please contact Nurse Assist at (800) 649-6800																																															