

<b>Manufacturer name:</b>	Sol-Millennium/Mckesson Medical Surgical												
<b>About the Recall:</b>	Cypress Medical Products, LLC (Cypress) has notified McKesson Medical Surgical Inc. (MMS) of an Urgent Product Recall regarding all lots of McKesson brand low dead space ("LDS") syringes manufactured by Sol-Millennium. This notice has been issued because these products were found by the FDA to be substantially different from SolMillennium's current cleared 510(k)s with a major change or modification to the device's intended use without a 510(k) clearance that remain outside of their control beyond a distributor level.												
<b>Products(s) numbers/models that are affected:</b>	<table border="1" data-bbox="443 957 1539 1108"> <thead> <tr> <th data-bbox="443 957 587 1003">MMS #</th> <th data-bbox="587 957 776 1003">MFG Catalog #</th> <th data-bbox="776 957 1425 1003">Description</th> <th data-bbox="1425 957 1539 1003">Affected Lot(s)</th> </tr> </thead> <tbody> <tr> <td data-bbox="443 1003 587 1056">1150041</td> <td data-bbox="587 1003 776 1056">16-SN3C251S</td> <td data-bbox="776 1003 1425 1056">SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX1" (50/BX 8BX/CS)</td> <td data-bbox="1425 1003 1539 1056">All Lots</td> </tr> <tr> <td data-bbox="443 1056 587 1108">1150042</td> <td data-bbox="587 1056 776 1108">16-SN3C2558S</td> <td data-bbox="776 1056 1425 1108">SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX5/8" (50/BX 8BX/CS)</td> <td data-bbox="1425 1056 1539 1108">All Lots</td> </tr> </tbody> </table>	MMS #	MFG Catalog #	Description	Affected Lot(s)	1150041	16-SN3C251S	SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX1" (50/BX 8BX/CS)	All Lots	1150042	16-SN3C2558S	SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX5/8" (50/BX 8BX/CS)	All Lots
MMS #	MFG Catalog #	Description	Affected Lot(s)										
1150041	16-SN3C251S	SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX1" (50/BX 8BX/CS)	All Lots										
1150042	16-SN3C2558S	SYRINGE/NDL, SAFETY PREVENT HT3CC 25GX5/8" (50/BX 8BX/CS)	All Lots										
<b>URL Link(s) to Manufacturer Notice/Web Site:</b>	<a data-bbox="414 1167 667 1199" href="http://www.solm.com">www.solm.com</a> <a data-bbox="414 1230 915 1262" href="mailto:MMSRecalls@McKesson.com">MMSRecalls@McKesson.com</a>												
<b>Actions to be taken by customer/Patient:</b>	Immediately discontinue use of any product matching the affected items and lot numbers listed above. For questions regarding this notification, please contact Sol-Millennium at (833) 637-0857.												