

Manufacturer name:	Convatec												
About the Recall:	<p>This notice has been issued because it has been identified that a number of cartons of 839002 Eakin Cohesive small seals with lot number 109357C553 contain individually blistered Eakin Cohesive small Seals which have been mis-labelled as 839005 Eakin Cohesive slim seals. In some cases, all 20 individually blistered seals inside the carton may be labelled as 839005. In other cases, some cartons may contain a mixture of individually blistered seals labelled as 839002 or 839005. Most cartons will contain individually blistered seals which are correctly labeled as 839002. Any mis-labelled individually blistered seals will not be evident until the carton has been opened. The product within the carton is the correct product, it is the labeling on the actual packaging that is mis-labelled.</p>												
Products(s) numbers/models that are affected:	<table border="1" data-bbox="505 1136 1555 1257"> <thead> <tr> <th data-bbox="505 1136 638 1188">MMS #</th> <th data-bbox="638 1136 868 1188">MFG Catalog #</th> <th data-bbox="868 1136 1313 1188">Description</th> <th data-bbox="1313 1136 1555 1188">Affected Lot(s)</th> </tr> </thead> <tbody> <tr> <td data-bbox="505 1188 638 1220">364592</td> <td data-bbox="638 1188 868 1220">839002</td> <td data-bbox="868 1188 1313 1220">SEAL, COHESIVE 2" SM (20/PK)</td> <td data-bbox="1313 1188 1555 1220">109357C553</td> </tr> <tr> <td data-bbox="505 1220 638 1257">697218</td> <td data-bbox="638 1220 868 1257">839002</td> <td data-bbox="868 1220 1313 1257">SEAL, COHESIVE 2" SM (20/BX)</td> <td data-bbox="1313 1220 1555 1257">109357C553</td> </tr> </tbody> </table>	MMS #	MFG Catalog #	Description	Affected Lot(s)	364592	839002	SEAL, COHESIVE 2" SM (20/PK)	109357C553	697218	839002	SEAL, COHESIVE 2" SM (20/BX)	109357C553
MMS #	MFG Catalog #	Description	Affected Lot(s)										
364592	839002	SEAL, COHESIVE 2" SM (20/PK)	109357C553										
697218	839002	SEAL, COHESIVE 2" SM (20/BX)	109357C553										
URL Link(s) to Manufacturer Notice/Web Site:	Convatec - Solutions for Management of Chronic Conditions												
Actions to be taken by customer/Patient:	<p>Immediately discontinue use of any product matching the affected item(s) and lot number(s) listed above. If you have no product matching the affected item(s) and lot number(s), no further action is needed</p>												