

<b>Manufacturer name:</b>	Unomedical												
<b>About the Recall:</b>	A Voluntary Medical Device Recall (removal) for a specific lot of the VariSoft™ Infusion Sets. These specific VariSoft™ Infusion Sets were packaged using incorrect packaging (boxes) branded "Trusteel™ Infusion Set". All components within the infusion set carton, including the Instructions for Use, are correct and safe to use.												
<b>Products(s) numbers/models that are affected:</b>	<table border="1" data-bbox="391 814 1479 915"> <thead> <tr> <th data-bbox="391 814 647 856">Description</th> <th data-bbox="647 814 805 856">SAP-Code</th> <th data-bbox="805 814 950 856">ICC-Code</th> <th data-bbox="950 814 1065 856">Lot#</th> <th data-bbox="1065 814 1312 856">Date of Manufacture</th> <th data-bbox="1312 814 1479 856">Expiry Date</th> </tr> </thead> <tbody> <tr> <td data-bbox="391 856 647 915">VariSoft™ Infusion Set</td> <td data-bbox="647 856 805 915">1726035</td> <td data-bbox="805 856 950 915">509953</td> <td data-bbox="950 856 1065 915">6005241</td> <td data-bbox="1065 856 1312 915">JAN-01-2024</td> <td data-bbox="1312 856 1479 915">DEC-01-2026</td> </tr> </tbody> </table>	Description	SAP-Code	ICC-Code	Lot#	Date of Manufacture	Expiry Date	VariSoft™ Infusion Set	1726035	509953	6005241	JAN-01-2024	DEC-01-2026
Description	SAP-Code	ICC-Code	Lot#	Date of Manufacture	Expiry Date								
VariSoft™ Infusion Set	1726035	509953	6005241	JAN-01-2024	DEC-01-2026								
<b>URL Link(s) to Manufacturer Notice/Web Site:</b>	<a data-bbox="370 961 691 995" href="http://www.convatec.com">www.convatec.com</a> <a data-bbox="370 1024 734 1058" href="http://www.infusion-set.com">www.infusion-set.com</a>												
<b>Actions to be taken by customer/Patient:</b>	<p data-bbox="370 1213 1500 1297">AdaptHealth has removed and quarantined all product that has been affected by this voluntary recall.</p> <p data-bbox="370 1327 1576 1453">For further information regarding this recall, you may contact the manufacturer directly by visiting <a data-bbox="899 1369 1211 1402" href="http://www.convatec.com">www.convatec.com</a> or calling 1-800-422-8811</p>												