

Manufacturer name:	Ambu Inc
About the Recall:	Ambu received 5 complaints concerning Ambu SPUR II deviating from the design with the manometer port being blocked rendering the manometer non-functional. As a result, users are unable to use the attached manometer to monitor the patient's airway pressure.
<b>B</b> 1 4 / \	

## Products(s) numbers/models that are affected:

MMS#	MFG Catalog #	Description	Affected Lot(s)	
770644	523211000	RESUSITATOR, SPUR II BG RSVR ADLT (12/CS)	1001113554	1001118764
557737	530212000	SPUR II, MEDI PORT TOD MASK PEDI BAG (12/CS)	1001113557	
533885	530213000	RESUSCITATOR, SPUR II PED TOD MASK (12/CS)	1001110299	1001118767
555665			1001113558	1001106634
1067666	530213030	SPUR II CHILD RESC W/MANOMETER(12/CS)	2000016409	
821459	530213031	RESUSCITATOR, SPUR II BG RESVRPED (12/CS)	1001113560	1001118769
555742	530214000	RESUSCITATOR, SPUR PED BG MEDIPRT /INF/TODD (12 AMBU)	1001106636	1001113561
544776	530613000	RESUSCITATOR, SPUR	1001110301	1001106637
544776		W/MASK PED	1001121455	1001118771
545840	531613000	RESUSCITATOR, SPUR II W/CORR TU PED MASK (6/CS)	1001118773	1001106640
			1001121460	
1205547	523611051E	RESUSCITATOR, MANUAL SPUR II ADLT MED MASK (6/CS)	2000015292	2000016401
1200047			2000015559	
876838	530213000B	BAG, RESUSCITATION PED W/MASK (12/CS)	1001113559	

	++ adapthealth
URL Link(s) to Manufacturer Notice/Web Site:	Ambu.com
Actions to be taken by customer/Patient:	Check your supply against the product list above, should you have any affected product evaluate the product for any possible defect and contact Ambu Inc. for product replacement. If you have any questions regarding this Urgent Medical Device Recall notice, please contact
	Ambu Inc. (800) 262-8462 for further information.