

Manufacturer name:	Integra Lif	eScie	ences						
About the Recall:	Wound & failures we products, the identificompletely packaging	Integra LifeSciences has issued a voluntary recall of the MediHoney Wound & Burn products. Reason for Device Correction Packaging failures were identified related to the MediHoney Wound & Burn products, which could lead to a breach in the sterile barriers. Some of the identified failures are as follows: 1. Tube foil induction seal is not completely sealed to the tube 2. Inadequate sealing of sterile barrier packaging 3. Shipping boxed did not adequately protect device during transportation 4. Tube twist-off cap failure.							
Products(s) numbers/models	Manufacturer Product # (Catalog #)	Issue #	Product Name (Description)	UDI Number	Lot # Expiration Date	Distribution Dates (MM/DD/YYYY)			
that are affected:	31515	1,2,3	MEDIHONEY TUBE - 1.5 FL OZ (44 ML) – W/APPLICATOR- STERILE	10381780486824	All unexpired lots	09/20/2021 to 03/28/2025			
			MEDITIONEN/TUDE OFFI OF			07144100044			

Manufacturer Product # (Catalog #)	Issue #	Product Name (Description)	UDI Number	Lot # Expiration Date	Distribution Dates (MM/DD/YYYY)
31515	1,2,3	MEDIHONEY TUBE - 1.5 FL OZ (44 ML) – W/APPLICATOR- STERILE	10381780486824	All unexpired lots	09/20/2021 to 03/28/2025
31535	1,2,3	MEDIHONEY TUBE - 3.5 FL OZ W/APPLICATOR - STERILE	10381780486831	All unexpired lots	07/14/2021 to 02/12/2025
31505	2, 3, 4	MEDIHONEY IN TUBE .5 FL OZ, TWISTOFF - STERILE	10381780486930	All unexpired lots	02/22/2022 to 03/26/2025
31612	2	MEDIHONEY HCS 8"X12" BURN DRESSING STERILE 1'S, 2/BOX X 5PK:10/CS	10381780486947	All unexpired lots	03/18/2024 to 03/19/2025
31620	2	MEDIHONEY HYDROGEL 2.4" X 2.4" STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471370	All unexpired lots	03/26/2024 to 03/25/2025
31622	2	MEDIHONEY HYDROGEL 2.4" X 2.4" SHEET STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471356	All unexpired lots	07/31/2023 to 07/30/2024
31640	2	MEDIHONEY HYDROGEL 4.3" X 4.3" STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471400	All unexpired lots	08/31/2023 to 03/24/2025
31644	2	MEDIHONEY HYDROGEL 4.3" X 4.3" SHEET STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471363	All unexpired lots	07/31/2023 to 07/01/2024
31720	2	MEDIHONEY HYDROGEL 2.8" X 2.8" ADHESIVE STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471417	All unexpired lots	04/03/2023 to 03/24/2025
31722	2	MEDIHONEY HYDROGEL 2.8" X 2.8" ADHESIVE SHEET STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471387	All unexpired lots	01/10/2024 to 01/27/2025
31738	2	MEDIHONEY HCS 3" X 8" STERILE 1'S, 10/BOX X 5PK:50/CS	10381780486961	All unexpired lots	02/06/2025 to 03/13/2025
31740	2	MEDIHONEY HYDROGEL 4.5"X 4.5" ADHESIVE STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471424	All unexpired lots	07/05/2024 to 03/21/2025
31744	2	MEDIHONEY HYDROGEL 4.5" X4.5" ADHESIVE SHEET STERILE 1'S, 10/BOX X 5PK:50/CS	10381780471394	All unexpired lots	12/15/2023 to 11/28/2024

URL Link(s) to Manufacturer Notice/Web Site:

www.integralife.com



Actions to be taken by customer/Patient:

Please check your supplies and reference the product listing provided above, should you have product please contact Integra LifeSciences for replacement of product.

Integra LifeSciences may be contacted Monday - Friday 8:00 AM - 8:00 PM EST at 1-800-654-2873, or via email custsvcnj@integralife.com